
Midmark M11 - v1.0.0
Total Cycles: 3127

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.1 F 0.0
2:00 190.6 F 0.0
4:00 212.0 F 5.5
6:00 228.7 F 11.1
8:00 229.7 F 7.3
10:00 243.3 F 13.3
12:00 242.1 F 11.7
14:00 240.5 F 10.6
16:00 254.4 F 17.5
18:00 266.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.8
2:00 271.4 F 28.1
3:00 272.7 F 28.9
4:00 272.0 F 28.3

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 219.9 F 0.0
10:00 219.5 F 0.0
15:00 228.0 F 0.0
20:00 229.1 F 0.0
25:00 229.2 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

04/19/2023 10:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____