

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3542

Name: RITTER2  
Sterilizer ID: V2313228  
04/18/2024 11:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.3 F 0.0  
2:00 159.3 F 2.5  
4:00 174.6 F 6.8  
6:00 203.2 F 11.8  
8:00 217.6 F 7.2  
10:00 228.8 F 12.7  
12:00 233.2 F 10.0  
14:00 235.9 F 8.7  
16:00 247.8 F 15.1  
18:00 259.9 F 21.9  
20:00 271.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 30.2  
1:00 271.6 F 30.3  
2:00 271.6 F 30.4  
3:00 271.6 F 30.5  
4:00 271.5 F 30.5  
  
Min 271.5 F 30.0  
Max 272.8 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 210.0 F 0.0  
10:00 215.7 F 0.0  
15:00 218.7 F 0.0  
20:00 219.9 F 0.0  
25:00 220.4 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 21:22  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:00:12  
CYCLE COMPLETE

04/18/2024 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_