
Midmark M11 - v1.0.0
Total Cycles: 5208

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2024 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.6 F 0.0
2:00 103.2 F 0.0
4:00 117.6 F 0.0
6:00 148.2 F 0.0
8:00 177.1 F 0.1
10:00 196.0 F 0.4
12:00 211.6 F 2.0
14:00 223.3 F 5.9
16:00 232.3 F 9.5
18:00 242.3 F 12.5
20:00 240.9 F 10.9
22:00 236.5 F 8.7
24:00 241.3 F 10.3
26:00 248.5 F 14.0
28:00 259.7 F 20.2
30:00 269.5 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.0 F 28.2
2:00 272.7 F 28.7
3:00 272.7 F 28.6
4:00 272.4 F 28.3

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 206.8 F 0.0
10:00 206.0 F 0.0
15:00 224.2 F 0.0
20:00 232.4 F 0.0
25:00 228.9 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:09:40
CYCLE COMPLETE

12/19/2024 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____