
Midmark M11 - v1.0.0
Total Cycles: 719

Name: MIDMARK
Sterilizer ID: V2178198
03/02/2021 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 160.4 F 2.5
4:00 174.4 F 6.4
6:00 200.5 F 11.9
8:00 218.0 F 7.0
10:00 232.0 F 13.1
12:00 238.1 F 10.4
14:00 235.7 F 8.6
16:00 251.3 F 16.1
18:00 264.2 F 23.5
20:00 271.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.0 F 28.8
2:00 272.1 F 28.8
3:00 272.4 F 28.8
4:00 272.6 F 29.0

Min 271.7 F 28.8
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 236.1 F 0.0
10:00 206.2 F 0.0
15:00 203.1 F 0.0
20:00 198.2 F 0.0
25:00 196.8 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:30
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

03/02/2021 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____