
Midmark M11 - v1.0.5
Total Cycles: 2951

Name: RITTER2
Sterilizer ID: V2313228
09/29/2023 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 172.4 F 3.0
4:00 191.3 F 7.3
6:00 208.6 F 11.9
8:00 225.4 F 6.8
10:00 236.7 F 11.6
12:00 235.0 F 8.8
14:00 240.4 F 10.8
16:00 246.1 F 13.5
18:00 257.5 F 19.7
20:00 267.8 F 26.1
22:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.1 F 29.1
2:00 272.5 F 29.4
3:00 272.0 F 29.6
4:00 271.5 F 29.3

Min 271.4 F 29.0
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 223.6 F 0.0
10:00 214.4 F 0.0
15:00 212.8 F 0.0
20:00 200.5 F 0.0
25:00 192.9 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

09/29/2023 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____