

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3347

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2024 03:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.0 F 0.0  
2:00 142.7 F 1.9  
4:00 163.9 F 5.5  
6:00 189.9 F 10.1  
8:00 213.8 F 6.0  
10:00 224.8 F 9.4  
12:00 232.1 F 9.5  
14:00 240.5 F 12.4  
16:00 240.9 F 11.3  
18:00 254.7 F 18.3  
20:00 266.9 F 25.9  
22:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.5  
1:00 272.2 F 29.4  
2:00 272.5 F 29.7  
3:00 272.6 F 30.1  
4:00 271.8 F 29.6

Min 271.7 F 29.4  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 230.3 F 0.0  
10:00 208.1 F 0.0  
15:00 205.4 F 0.0  
20:00 201.6 F 0.0  
25:00 198.3 F 0.0  
30:00 196.5 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 22:05  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:25  
CYCLE COMPLETE

02/01/2024 04:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_