

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4689

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2024 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.9 F 0.0  
2:00 165.6 F 2.9  
4:00 188.5 F 7.0  
6:00 212.7 F 11.9  
8:00 219.2 F 6.8  
10:00 234.5 F 12.1  
12:00 233.8 F 9.2  
14:00 237.7 F 9.2  
16:00 247.3 F 14.1  
18:00 258.6 F 20.6  
20:00 270.0 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.5 F 29.4  
2:00 271.8 F 28.8  
3:00 271.7 F 29.1  
4:00 272.5 F 29.1

Min 271.0 F 28.5  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 232.8 F 0.0  
10:00 216.9 F 0.0  
15:00 230.7 F 0.0  
20:00 223.5 F 0.0  
25:00 223.0 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 20:48  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

06/28/2024 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_