
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003576
14:01 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:12
STERILIZING 3:07
136.0 C 220kPa 3:07
136.4 C 223kPa 3:37
136.3 C 221kPa 4:07
136.4 C 224kPa 4:37
136.4 C 225kPa 5:07
136.3 C 224kPa 5:37
136.4 C 224kPa 6:07
136.2 C 222kPa 6:37
136.2 C 222kPa 6:37
VENTING 6:37
AIR DRYING 7:26
CYCLE COMPLETE 7:27

Digital Signature #
4E8A255B903C67F5
