
Midmark M11 - v1.0.0
Total Cycles: 3189

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2023 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.6 F 0.0
2:00 141.8 F 0.0
4:00 167.3 F 1.4
6:00 191.8 F 5.4
8:00 211.6 F 9.7
10:00 223.0 F 8.1
12:00 231.4 F 9.9
14:00 233.7 F 8.6
16:00 243.4 F 12.9
18:00 241.9 F 11.6
20:00 255.1 F 18.3
22:00 266.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.3
2:00 271.8 F 28.6
3:00 272.1 F 29.2
4:00 272.5 F 29.2

Min 271.5 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 219.5 F 0.0
10:00 209.9 F 0.0
15:00 210.5 F 0.0
20:00 217.8 F 0.0
25:00 225.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

05/02/2023 03:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____