
Midmark M11 - v1.0.5
Total Cycles: 3256

Name: RITTER2
Sterilizer ID: V2313228
12/27/2023 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.7 F 0.0
2:00 114.9 F 0.4
4:00 138.1 F 2.6
6:00 165.8 F 6.3
8:00 189.8 F 10.9
10:00 220.4 F 5.1
12:00 230.4 F 10.0
14:00 231.8 F 8.0
16:00 241.0 F 11.7
18:00 238.2 F 9.9
20:00 251.6 F 16.5
22:00 263.4 F 23.4
24:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.7 F 29.4
2:00 272.8 F 29.8
3:00 272.1 F 29.9
4:00 271.9 F 29.3

Min 271.5 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 237.3 F 0.0
10:00 201.1 F 0.0
15:00 185.9 F 0.0
20:00 180.9 F 0.0
25:00 179.8 F 0.0
30:00 179.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

12/27/2023 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____