
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.5uS / 0.9ppm
CYCLE NUMBER 000599
14:40 25/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:47
STERILIZING 3:59
136.0 C 225kPa 3:59
136.2 C 229kPa 4:29
136.4 C 229kPa 4:59
136.4 C 230kPa 5:29
136.4 C 232kPa 5:59
136.2 C 230kPa 6:29
136.4 C 230kPa 6:59
136.4 C 229kPa 7:29
136.4 C 229kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 23:21

Digital Signature #
65224B045CFB039D
