

-----  
Midmark M11 - v1.0.5  
Total Cycles: 39

Name: RITTER2  
Sterilizer ID: V2313228  
05/20/2021 01:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
2:00 182.2 F 3.5  
4:00 187.1 F 7.5  
6:00 207.2 F 11.7  
8:00 220.0 F 5.9  
10:00 230.9 F 10.5  
12:00 233.6 F 8.7  
14:00 243.9 F 12.9  
16:00 245.6 F 13.3  
18:00 257.0 F 19.6  
20:00 267.9 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 271.8 F 29.3  
2:00 272.7 F 29.4  
3:00 271.8 F 28.9  
4:00 271.3 F 28.9

Min 271.3 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
5:00 228.7 F 0.0  
10:00 223.6 F 0.0  
15:00 219.6 F 0.0  
20:00 218.1 F 0.0  
25:00 212.2 F 0.0  
30:00 202.1 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 21:16  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

05/20/2021 02:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_