
Midmark M11 - v1.0.5
Total Cycles: 1422

Name: RITTER2
Sterilizer ID: V2313228
08/29/2022 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 74.6 F 0.0
4:00 110.9 F 0.0
6:00 102.4 F 0.0
8:00 133.6 F 1.3
10:00 156.2 F 2.9
12:00 178.6 F 5.0
14:00 203.4 F 7.8
16:00 216.4 F 11.4
18:00 222.5 F 5.9
20:00 233.2 F 10.6
22:00 233.7 F 8.4
24:00 243.4 F 13.0
26:00 241.9 F 11.7
28:00 253.8 F 17.7
30:00 264.3 F 23.9
32:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.5
2:00 271.5 F 28.9
3:00 272.9 F 29.5
4:00 271.6 F 28.8

Min 271.5 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 202.5 F 0.0
10:00 207.6 F 0.0
15:00 217.8 F 0.0
20:00 217.6 F 0.0
25:00 218.0 F 0.0
30:00 217.0 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:09
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

08/29/2022 08:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____