
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007821
10:03 04/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:22
STERILIZING 4:22
136.4 C 223kPa 4:22
136.2 C 223kPa 4:52
136.2 C 220kPa 5:22
136.4 C 222kPa 5:52
136.4 C 221kPa 6:22
136.1 C 222kPa 6:52
136.4 C 224kPa 7:22
136.0 C 221kPa 7:52
136.0 C 221kPa 7:52
VENTING 7:52
AIR DRYING 8:42
CYCLE COMPLETE 38:43

Digital Signature #
96C01037C7C256E2
