
Midmark M11 - v1.0.5
Total Cycles: 1054

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 187.2 F 3.9
4:00 209.3 F 9.6
6:00 224.5 F 7.1
8:00 234.9 F 11.7
10:00 236.3 F 10.0
12:00 236.4 F 8.9
14:00 250.7 F 16.3
16:00 262.8 F 23.3
18:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 30.0
2:00 272.3 F 29.5
3:00 271.3 F 29.2
4:00 272.3 F 30.1

Min 271.3 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 226.4 F 0.0
10:00 227.2 F 0.0
15:00 227.8 F 0.0
20:00 225.8 F 0.0
25:00 225.1 F 0.0
30:00 223.5 F 0.0

Drying Complete

FILLING: 3:53
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

05/02/2022 11:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____