
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001855
15:00 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:35
136.0 C 222kPa 3:35
136.0 C 222kPa 4:05
136.4 C 226kPa 4:35
136.4 C 226kPa 5:05
136.4 C 224kPa 5:35
136.4 C 226kPa 6:05
136.3 C 224kPa 6:35
136.3 C 224kPa 7:05
136.3 C 224kPa 7:05
VENTING 7:05
AIR DRYING 8:00
CYCLE COMPLETE 8:02

Digital Signature #
393463A84AA506ED
