
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005405
10:56 21/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:46
PRESSURIZING 1:50
STERILIZING 2:54
136.5 C 220kPa 2:54
136.3 C 223kPa 3:25
136.2 C 218kPa 3:55
136.2 C 222kPa 4:25
136.2 C 223kPa 4:55
136.3 C 222kPa 5:25
136.4 C 224kPa 5:55
136.1 C 223kPa 6:25
136.1 C 223kPa 6:25
VENTING 6:25
AIR DRYING 7:14
CYCLE COMPLETE 8:35

Digital Signature #
1CCE20C27F13CDD7
