
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001023
15:48 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:38
STERILIZING 3:41
136.0 C 222kPa 3:41
136.0 C 222kPa 4:11
136.4 C 225kPa 4:41
136.4 C 226kPa 5:11
136.1 C 224kPa 5:41
136.2 C 224kPa 6:11
136.4 C 225kPa 6:41
136.2 C 224kPa 7:11
136.2 C 224kPa 7:11
VENTING 7:11
AIR DRYING 8:02
CYCLE COMPLETE 14:29

Digital Signature #
5BED51FB8F81623E
