
Midmark M11 - v1.0.0
Total Cycles: 4717

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2024 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.0 F 0.0
2:00 122.3 F 1.3
4:00 128.7 F 4.0
6:00 156.6 F 7.2
8:00 188.3 F 11.1
10:00 211.2 F 4.1
12:00 220.9 F 8.3
14:00 228.5 F 12.6
16:00 231.8 F 8.4
18:00 242.5 F 13.5
20:00 241.3 F 11.4
22:00 252.8 F 17.2
24:00 263.4 F 23.2
26:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.3 F 29.5
2:00 272.2 F 29.3
3:00 272.4 F 29.4
4:00 271.5 F 29.0

Min 271.0 F 28.6
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 205.2 F 0.0
10:00 191.0 F 0.0
15:00 188.9 F 0.0
20:00 198.3 F 0.0
25:00 216.3 F 0.0
30:00 228.9 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 26:24
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:03
CYCLE COMPLETE

07/11/2024 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____