
Midmark M11 - v1.0.0
Total Cycles: 1917

Name: RITTER 1
Sterilizer ID: V2178198
05/09/2022 07:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 72.8 F 0.0
4:00 81.5 F 0.0
6:00 119.2 F 0.3
8:00 151.6 F 1.9
10:00 177.7 F 4.2
12:00 196.0 F 7.1
14:00 214.1 F 10.4
16:00 223.0 F 7.3
18:00 228.1 F 7.1
20:00 238.1 F 11.2
22:00 232.9 F 7.3
24:00 246.2 F 13.2
26:00 243.4 F 11.5
28:00 255.8 F 18.0
30:00 266.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.8
1:00 272.5 F 28.4
2:00 271.2 F 27.7
3:00 272.5 F 28.6
4:00 271.8 F 27.9

Min 271.1 F 27.7
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 225.0 F 0.0
10:00 186.1 F 0.0
15:00 193.5 F 0.0
20:00 198.8 F 0.0
25:00 247.0 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:36
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

05/09/2022 08:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____