
Midmark M11 - v1.0.5
Total Cycles: 3126

Name: RITTER2
Sterilizer ID: V2313228
11/15/2023 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 153.3 F 1.2
4:00 160.8 F 4.3
6:00 187.7 F 8.4
8:00 203.3 F 13.3
10:00 221.4 F 7.6
12:00 233.0 F 12.8
14:00 235.5 F 10.0
16:00 236.6 F 8.9
18:00 248.3 F 15.0
20:00 260.1 F 21.6
22:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.9 F 29.7
2:00 271.5 F 29.0
3:00 272.1 F 29.8
4:00 272.6 F 29.5

Min 271.0 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 210.1 F 0.0
10:00 204.4 F 0.0
15:00 199.4 F 0.0
20:00 197.1 F 0.0
25:00 196.0 F 0.0
30:00 197.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

11/15/2023 12:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____