
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002073
11:57 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:58
STERILIZING 3:02
136.5 C 226kPa 3:02
136.3 C 225kPa 3:32
136.3 C 224kPa 4:02
136.2 C 224kPa 4:32
136.4 C 224kPa 5:02
136.4 C 226kPa 5:32
136.3 C 224kPa 6:02
136.3 C 225kPa 6:32
136.3 C 225kPa 6:32
VENTING 6:32
AIR DRYING 7:23
CYCLE COMPLETE 17:02

Digital Signature #
3FD91106290EB17A
