
Midmark M11 - v1.0.5
Total Cycles: 2823

Name: RITTER2
Sterilizer ID: V2313228
08/31/2023 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.5 F 0.0
2:00 149.9 F 0.0
4:00 177.0 F 0.7
6:00 190.9 F 4.1
8:00 207.5 F 8.0
10:00 222.0 F 12.6
12:00 228.0 F 8.5
14:00 239.4 F 12.4
16:00 241.6 F 11.9
18:00 240.9 F 10.9
20:00 253.4 F 17.2
22:00 264.0 F 23.5
24:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.6 F 29.8
2:00 272.0 F 29.0
3:00 272.8 F 29.6
4:00 271.8 F 29.3

Min 270.9 F 28.4
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 220.2 F 0.0
10:00 194.8 F 0.0
15:00 192.0 F 0.0
20:00 188.0 F 0.0
25:00 185.3 F 0.0
30:00 183.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:33
CYCLE COMPLETE

08/31/2023 12:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____