

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.4uS / 0.8ppm  
CYCLE NUMBER 000598  
13:26 25/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:47  
STERILIZING 3:55  
136.0 C 226kPa 3:55  
136.4 C 232kPa 4:25  
136.0 C 229kPa 4:55  
136.2 C 230kPa 5:25  
136.4 C 232kPa 5:55  
136.4 C 232kPa 6:25  
136.1 C 229kPa 6:55  
136.2 C 230kPa 7:25  
136.2 C 230kPa 7:25  
VENTING 7:25  
AIR DRYING 8:16  
CYCLE COMPLETE 8:23  
-----

Digital Signature #  
2B511B1733A679E3  
-----