

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1293

Name: RITTER2  
Sterilizer ID: V2313228  
07/13/2022 03:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
2:00 176.0 F 0.1  
4:00 200.6 F 2.3  
6:00 210.9 F 7.2  
8:00 224.0 F 12.7  
10:00 231.6 F 9.2  
12:00 237.2 F 10.1  
14:00 245.0 F 13.1  
16:00 245.3 F 13.3  
18:00 258.4 F 20.3  
20:00 269.9 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.3 F 29.2  
2:00 271.9 F 29.4  
3:00 272.0 F 29.0  
4:00 272.1 F 29.6  
  
Min 271.5 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.8 F 0.0  
5:00 223.7 F 0.0  
10:00 235.6 F 0.0  
15:00 234.5 F 0.0  
20:00 228.7 F 0.0  
25:00 218.3 F 0.0  
30:00 216.9 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 20:53  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

07/13/2022 04:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_