
Midmark M11 - v1.0.5
Total Cycles: 393

Name: RITTER2
Sterilizer ID: V2313228
09/27/2021 06:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 82.2 F 0.0
4:00 126.6 F 0.1
6:00 115.3 F 0.2
8:00 138.8 F 1.6
10:00 166.9 F 4.5
12:00 189.6 F 7.9
14:00 218.8 F 12.5
16:00 226.6 F 6.6
18:00 235.6 F 10.9
20:00 232.8 F 7.5
22:00 243.2 F 13.3
24:00 242.1 F 12.0
26:00 254.9 F 18.5
28:00 265.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.8 F 29.1
2:00 272.1 F 29.5
3:00 271.7 F 29.6
4:00 271.4 F 28.9

Min 271.2 F 28.8
Max 272.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 186.1 F 0.0
10:00 181.5 F 0.0
15:00 176.5 F 0.0
20:00 172.4 F 0.0
25:00 212.8 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:39
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

09/27/2021 07:41 AM
ACCEPTED: X
REJECTED:
Operator: U01