
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001656
10:11 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:39
STERILIZING 3:46
136.0 C 220kPa 3:46
136.1 C 221kPa 4:16
136.2 C 222kPa 4:46
136.3 C 221kPa 5:16
136.2 C 222kPa 5:46
136.3 C 223kPa 6:16
136.4 C 224kPa 6:46
136.4 C 224kPa 7:16
136.4 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:08
CYCLE COMPLETE 10:21

Digital Signature #
C1EBA0357AF79B36
