
Midmark M11 - v1.0.0
Total Cycles: 964

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 78.4 F 0.0
4:00 92.8 F 0.0
6:00 121.7 F 0.0
8:00 150.0 F 0.1
10:00 177.6 F 0.2
12:00 196.1 F 0.9
14:00 210.5 F 3.7
16:00 217.5 F 6.0
18:00 226.2 F 8.8
20:00 235.5 F 12.4
22:00 233.3 F 8.5
24:00 244.6 F 12.8
26:00 243.2 F 12.0
28:00 241.3 F 10.8
30:00 254.1 F 17.3
32:00 265.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.2 F 28.7
2:00 272.3 F 28.4
3:00 271.7 F 28.3
4:00 272.7 F 28.8

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 164.8 F 0.0
5:00 216.3 F 0.0
10:00 232.3 F 0.0
15:00 233.2 F 0.0
20:00 219.2 F 0.0
25:00 206.9 F 0.0
30:00 208.7 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:46
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:11:55
CYCLE COMPLETE

07/06/2021 07:56 AM
ACCEPTED: X
REJECTED:
Operator: U01