
Midmark M11 - v1.0.5
Total Cycles: 3771

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.3 F 0.0
2:00 157.1 F 2.3
4:00 172.8 F 5.9
6:00 202.5 F 10.6
8:00 218.1 F 5.6
10:00 226.2 F 10.7
12:00 230.5 F 7.8
14:00 240.6 F 11.8
16:00 244.5 F 13.2
18:00 256.7 F 19.8
20:00 268.5 F 26.8
22:00 272.2 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.8
1:00 271.7 F 29.8
2:00 271.8 F 29.7
3:00 271.8 F 29.7
4:00 272.1 F 29.8

Min 271.5 F 29.7
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 232.5 F 0.0
10:00 192.9 F 0.0
15:00 189.6 F 0.0
20:00 185.2 F 0.0
25:00 184.9 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

07/11/2024 04:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____