
Midmark M11 - v1.0.5
Total Cycles: 1588

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 180.7 F 3.5
4:00 188.1 F 8.0
6:00 211.1 F 13.5
8:00 225.8 F 8.8
10:00 235.6 F 10.3
12:00 241.5 F 12.7
14:00 241.8 F 11.9
16:00 255.0 F 18.8
18:00 266.6 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.8 F 29.9
2:00 271.6 F 29.1
3:00 272.4 F 30.1
4:00 272.2 F 29.6

Min 271.3 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 215.1 F 0.0
10:00 227.5 F 0.0
15:00 229.2 F 0.0
20:00 229.3 F 0.0
25:00 227.0 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

10/12/2022 04:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____