

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1906

Name: RITTER 1  
Sterilizer ID: V2178198  
05/04/2022 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.3 F 0.0  
2:00 194.6 F 2.0  
4:00 190.0 F 6.2  
6:00 210.4 F 11.0  
8:00 222.2 F 5.7  
10:00 233.5 F 11.2  
12:00 234.6 F 8.0  
14:00 242.9 F 11.3  
16:00 244.9 F 12.8  
18:00 257.5 F 19.4  
20:00 268.6 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.3  
2:00 272.1 F 28.7  
3:00 271.5 F 28.6  
4:00 272.7 F 29.2  
  
Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 212.2 F 0.0  
10:00 223.8 F 0.0  
15:00 232.3 F 0.0  
20:00 232.5 F 0.0  
25:00 227.7 F 0.0  
30:00 225.7 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:14  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:27  
CYCLE COMPLETE

05/04/2022 09:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_