

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3316

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2023 03:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.9 F 0.0  
2:00 150.0 F 0.0  
4:00 174.0 F 2.1  
6:00 199.4 F 6.5  
8:00 219.8 F 11.3  
10:00 223.9 F 6.5  
12:00 236.0 F 11.8  
14:00 235.8 F 9.2  
16:00 238.7 F 9.5  
18:00 248.1 F 14.3  
20:00 260.7 F 21.2  
22:00 272.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 28.9  
2:00 271.5 F 28.3  
3:00 272.6 F 29.0  
4:00 271.8 F 28.3

Min 271.4 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 218.5 F 0.0  
10:00 221.5 F 0.0  
15:00 220.2 F 0.0  
20:00 221.3 F 0.0  
25:00 221.7 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 22:32  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

05/31/2023 04:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_