
Midmark M11 - v1.0.5
Total Cycles: 2453

Name: RITTER2
Sterilizer ID: V2313228
06/01/2023 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.5 F 0.0
2:00 194.3 F 4.3
4:00 210.7 F 10.2
6:00 222.9 F 6.2
8:00 236.3 F 12.1
10:00 237.8 F 10.4
12:00 237.3 F 9.4
14:00 252.1 F 16.7
16:00 264.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.3
2:00 271.7 F 29.0
3:00 272.7 F 29.3
4:00 271.5 F 28.7

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 212.6 F 0.0
10:00 244.8 F 0.0
15:00 236.5 F 0.0
20:00 226.5 F 0.0
25:00 222.7 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 17:50
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

06/01/2023 01:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____