
Midmark M11 - v1.0.5
Total Cycles: 3429

Name: RITTER2
Sterilizer ID: V2313228
03/01/2024 04:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 153.5 F 2.0
4:00 177.9 F 5.7
6:00 198.5 F 10.4
8:00 219.6 F 5.7
10:00 225.1 F 9.5
12:00 235.0 F 9.2
14:00 242.1 F 12.3
16:00 241.5 F 11.2
18:00 254.6 F 17.9
20:00 266.1 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.2 F 29.3
2:00 271.2 F 28.8
3:00 272.1 F 29.4
4:00 271.3 F 28.7

Min 271.0 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 219.9 F 0.0
10:00 210.9 F 0.0
15:00 191.7 F 0.0
20:00 185.3 F 0.0
25:00 183.5 F 0.0
30:00 182.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:14
CYCLE COMPLETE

03/01/2024 05:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____