
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008396
7:20 11/03/2025

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:28
PRESSURIZING 1:32
STERILIZING 2:32
136.4 C 220kPa 2:32
136.4 C 223kPa 3:02
136.4 C 222kPa 3:32
136.4 C 221kPa 4:02
136.4 C 219kPa 4:32
136.2 C 219kPa 5:02
136.3 C 219kPa 5:32
136.2 C 219kPa 6:02
136.2 C 219kPa 6:02
VENTING 6:03
AIR DRYING 6:50
CYCLE COMPLETE 33:56

Digital Signature #
682838230F665762
