
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001829
11:50 21/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:24
STERILIZING 4:36
136.1 C 220kPa 4:36
136.2 C 223kPa 5:06
136.4 C 223kPa 5:36
136.2 C 223kPa 6:06
136.2 C 222kPa 6:36
136.5 C 224kPa 7:06
136.2 C 223kPa 7:36
136.3 C 222kPa 8:06
136.3 C 222kPa 8:06
VENTING 8:06
AIR DRYING 8:59
CYCLE COMPLETE 9:00

Digital Signature #
9490671FA2D0C70B
