

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003179  
10:28 17/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:06  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
136.0 C 221kPa 4:22  
136.4 C 221kPa 4:52  
136.4 C 223kPa 5:22  
136.4 C 224kPa 5:52  
136.4 C 224kPa 6:22  
136.2 C 221kPa 6:52  
136.3 C 221kPa 7:22  
136.3 C 221kPa 7:22  
VENTING 7:22  
AIR DRYING 8:14  
CYCLE COMPLETE 8:39

-----  
Digital Signature #  
03100564CAC71127  
-----