
Midmark M11 - v1.0.5
Total Cycles: 1838

Name: RITTER2
Sterilizer ID: V2313228
12/14/2022 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.6 F 0.0
2:00 193.5 F 3.7
4:00 200.5 F 8.8
6:00 224.9 F 8.0
8:00 230.3 F 10.1
10:00 232.5 F 7.9
12:00 241.1 F 11.1
14:00 246.8 F 14.1
16:00 260.0 F 21.3
18:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.5 F 29.7
2:00 271.5 F 28.9
3:00 272.5 F 29.6
4:00 271.6 F 28.9

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 225.9 F 0.0
10:00 240.4 F 0.0
15:00 239.6 F 0.0
20:00 227.7 F 0.0
25:00 223.1 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

12/14/2022 12:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____