
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001589
13:29 11/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 5:46
STERILIZING 7:01
136.0 C 220kPa 7:01
136.3 C 217kPa 7:32
136.1 C 220kPa 8:02
136.4 C 221kPa 8:32
136.4 C 223kPa 9:02
136.4 C 221kPa 9:32
136.4 C 223kPa 10:02
136.5 C 223kPa 10:31
VENTING 10:32
AIR DRYING 11:34
CYCLE COMPLETE 14:00

Digital Signature #
6C0B19602E90C43C
