
Midmark M11 - v1.0.5
Total Cycles: 2214

Name: RITTER2
Sterilizer ID: V2313228
04/04/2023 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 186.2 F 3.4
4:00 196.7 F 8.5
6:00 215.9 F 9.3
8:00 230.4 F 10.1
10:00 232.7 F 8.1
12:00 240.7 F 11.0
14:00 246.9 F 14.1
16:00 259.9 F 21.4
18:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.1 F 29.4
2:00 271.7 F 29.5
3:00 272.0 F 29.3
4:00 271.5 F 29.4

Min 271.3 F 28.9
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 226.4 F 0.0
10:00 243.9 F 0.0
15:00 235.8 F 0.0
20:00 222.5 F 0.0
25:00 222.0 F 0.0
30:00 223.4 F 0.0

Drying Complete

FILLING: 1:53
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:39
CYCLE COMPLETE

04/04/2023 12:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____