

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000359  
13:18 20/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:33  
PRESSURIZING 1:40  
STERILIZING 2:39  
136.0 C 220kPa 2:39  
136.4 C 229kPa 3:09  
136.4 C 232kPa 3:39  
136.2 C 231kPa 4:09  
136.4 C 232kPa 4:39  
136.3 C 233kPa 5:09  
136.2 C 230kPa 5:39  
136.3 C 232kPa 6:09  
136.3 C 232kPa 6:09  
VENTING 6:09  
AIR DRYING 7:02  
CYCLE COMPLETE 7:19  
-----

Digital Signature #  
9F8E057C91D1EBCC  
-----