
Midmark M11 - v1.0.0
Total Cycles: 1978

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.3 F 0.0
2:00 122.2 F 0.3
4:00 133.2 F 1.7
6:00 162.8 F 4.9
8:00 189.8 F 9.2
10:00 220.5 F 8.1
12:00 223.5 F 7.5
14:00 233.5 F 11.9
16:00 232.5 F 7.9
18:00 243.6 F 13.4
20:00 241.6 F 11.3
22:00 253.6 F 17.5
24:00 264.5 F 24.0
26:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.4
2:00 271.7 F 28.7
3:00 272.1 F 29.3
4:00 272.4 F 29.2

Min 271.3 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 205.7 F 0.0
10:00 208.6 F 0.0
15:00 210.6 F 0.0
20:00 213.7 F 0.0
25:00 214.8 F 0.0
30:00 218.2 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 26:00
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:05
CYCLE COMPLETE

05/31/2022 10:22 AM
ACCEPTED: X
REJECTED:
Operator: U02