
Midmark M11 - v1.0.5
Total Cycles: 4105

Name: RITTER2
Sterilizer ID: V2313228
11/11/2024 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.7 F 0.0
2:00 68.0 F 0.0
4:00 75.5 F 0.0
6:00 95.7 F 0.1
8:00 129.4 F 0.4
10:00 164.3 F 0.5
12:00 189.6 F 1.9
14:00 206.1 F 4.9
16:00 215.7 F 8.2
18:00 221.8 F 12.0
20:00 228.1 F 7.6
22:00 239.5 F 12.3
24:00 237.9 F 10.2
26:00 237.2 F 9.2
28:00 249.0 F 15.2
30:00 260.1 F 21.5
32:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.1 F 29.1
2:00 272.6 F 29.5
3:00 272.2 F 29.6
4:00 271.5 F 28.9

Min 271.5 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 229.4 F 0.0
10:00 212.3 F 0.0
15:00 195.6 F 0.0
20:00 190.0 F 0.0
25:00 188.8 F 0.0
30:00 191.5 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:49
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:11:09
CYCLE COMPLETE

11/11/2024 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____