
Midmark M11 - v1.0.0
Total Cycles: 692

Name: MIDMARK
Sterilizer ID: V2178198
02/23/2021 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.2 F 0.0
2:00 92.5 F 0.0
4:00 118.6 F 0.1
6:00 147.3 F 0.2
8:00 184.4 F 2.2
10:00 203.3 F 5.8
12:00 215.5 F 9.5
14:00 234.4 F 10.5
16:00 234.3 F 9.8
18:00 235.9 F 8.6
20:00 245.0 F 13.3
22:00 243.1 F 11.9
24:00 256.2 F 18.7
26:00 267.6 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.5 F 28.3
2:00 271.7 F 28.4
3:00 272.0 F 28.6
4:00 272.2 F 28.9

Min 271.3 F 28.3
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 231.6 F 0.0
10:00 224.8 F 0.0
15:00 233.9 F 0.0
20:00 187.1 F 0.0
25:00 176.2 F 0.0
30:00 174.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 27:10
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:05:00
CYCLE COMPLETE

02/23/2021 03:52 PM
ACCEPTED: X
REJECTED:
Operator: U01