
Midmark M11 - v1.0.5
Total Cycles: 65

Name: RITTER2
Sterilizer ID: V2313228
06/01/2021 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.9 F 0.0
2:00 129.0 F 1.9
4:00 148.6 F 5.3
6:00 183.5 F 10.1
8:00 202.6 F 4.7
10:00 219.4 F 9.6
12:00 227.1 F 8.4
14:00 238.3 F 12.9
16:00 239.4 F 11.7
18:00 254.0 F 18.6
20:00 266.0 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.1
1:00 272.4 F 30.3
2:00 272.9 F 30.6
3:00 272.7 F 30.5
4:00 272.7 F 30.5

Min 271.8 F 29.9
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.6 F 0.0
5:00 240.7 F 0.0
10:00 217.4 F 0.0
15:00 223.5 F 0.0
20:00 213.1 F 0.0
25:00 220.2 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:28
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

06/01/2021 03:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____