
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001527
10:57 10/22/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:29
PRESSURIZING 6:14
STERILIZING 7:50
136.0 C 220kPa 7:50
136.0 C 218kPa 8:20
136.4 C 221kPa 8:50
136.2 C 221kPa 9:20
136.4 C 218kPa 9:50
136.4 C 222kPa 10:20
136.4 C 223kPa 10:50
136.3 C 221kPa 11:20
VENTING 11:20
AIR DRYING 12:22
CYCLE COMPLETE 20:56

Digital Signature #
5654547DC5A5EA8E
