
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002150
12:55 16/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:03
136.0 C 220kPa 3:03
136.1 C 221kPa 3:33
136.3 C 223kPa 4:03
136.2 C 223kPa 4:33
136.2 C 223kPa 5:03
136.3 C 223kPa 5:33
136.4 C 223kPa 6:03
136.4 C 225kPa 6:33
136.4 C 225kPa 6:33
VENTING 6:33
AIR DRYING 7:25
CYCLE COMPLETE 37:26

Digital Signature #
5EB6F961532B6C72
