
Midmark M11 - v1.0.0
Total Cycles: 1035

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 150.1 F 1.0
4:00 165.5 F 3.4
6:00 187.8 F 6.7
8:00 207.0 F 10.9
10:00 219.2 F 3.9
12:00 229.2 F 8.5
14:00 238.6 F 13.1
16:00 238.6 F 10.1
18:00 237.1 F 8.8
20:00 248.3 F 14.1
22:00 259.3 F 20.2
24:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.0
3:00 272.2 F 28.7
4:00 272.3 F 28.5

Min 271.2 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 229.4 F 0.0
10:00 217.7 F 0.0
15:00 226.0 F 0.0
20:00 229.7 F 0.0
25:00 215.2 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

07/21/2021 09:12 AM
ACCEPTED: X
REJECTED:
Operator: U01