
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008142
9:20 04/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:03
STERILIZING 4:12
136.4 C 222kPa 4:12
136.1 C 217kPa 4:42
136.4 C 224kPa 5:12
136.2 C 219kPa 5:42
136.2 C 221kPa 6:12
136.2 C 222kPa 6:42
136.3 C 221kPa 7:12
136.4 C 220kPa 7:42
136.4 C 220kPa 7:42
VENTING 7:42
AIR DRYING 8:35
CYCLE COMPLETE 16:06

Digital Signature #
422EBEE74571ACBA
