
Midmark M11 - v1.0.5
Total Cycles: 3642

Name: RITTER2
Sterilizer ID: V2313228
05/24/2024 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.4 F 0.0
2:00 133.4 F 1.2
4:00 141.4 F 3.4
6:00 166.9 F 6.6
8:00 189.7 F 10.6
10:00 226.2 F 6.4
12:00 224.8 F 8.8
14:00 238.9 F 13.3
16:00 232.3 F 9.9
18:00 236.4 F 9.3
20:00 245.4 F 13.6
22:00 256.3 F 19.6
24:00 267.1 F 26.0
26:00 271.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.8
1:00 272.4 F 30.3
2:00 272.2 F 30.6
3:00 271.4 F 30.0
4:00 273.3 F 30.6

Min 271.0 F 29.7
Max 273.6 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 208.0 F 0.0
10:00 186.4 F 0.0
15:00 181.9 F 0.0
20:00 177.9 F 0.0
25:00 175.7 F 0.0
30:00 175.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:11
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:44
CYCLE COMPLETE

05/24/2024 11:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____